Surcharge for changing the temperature range of TA and WiFi sensors up to +600 °C

code: SLU119



Temperature calibration certificate issued by an accredited calibration laboratory.
Surcharge for hardware modification of the standard temperature measurement range up to +600°C.
Only for Web Sensor 2.0 transmitters TA4611, TA4621, TA3465 or wireless WiFi sensors W0711 and W0741.

Devices TA4611, TA4621 and TA3465
Range: -200 °C to +600 °C
Accuracy: ±0.2 °C in the range up to +100 °C; ±0.2% of the measured value in the range above +100 °C

Devices W0711 and W0741
Range: -150 °C to +600 °C
Accuracy: ±0.2 °C in the range up to +100 °C; ±0.2% of the measured value in the range above +100 °C

This service is normally available only with the purchase of a new device. To modify the measurement range of a previously purchased device, please contact our service department.

Technical data

CO2 SENSOR	
Measuring range	0 to 10 000 ppm
Accuracy	±(110ppm +5% from reading) at 25°C and 1013hPa
Resolution	1 ppm